



श्री चित्रा तिरुनाल आयुर्विज्ञान और प्रौद्योगिकी संस्थान, तिरुवनंतपुरम- 11, केरल
Sree Chitra Tirunal Institute for Medical Sciences and Technology, Thiruvananthapuram- 11, Kerala
(भारत सरकार के अधीन राष्ट्रीय महत्व संस्थान)
(An Institute of National Importance under Government of India)
टेलीफोन नं./Telephone No. 0471-2443152 फाक्स/Fax 0471-2446433 / 2550728
ई-मेल/E-mail :sct@sctimst.ac.in वेबसाइट/ Website : www.sctimst.ac.in

CORRIGENDUM - 4 dtd. 24.12.2019

TENDER NO. SCT/H/PMSSY/I/2019-20/16

| SCHEDULE I | | |
|---|---|---|
| SINGLE PLANE CATHLAB WITH CATH RECORDER FOR CARDIOLOGY | | |
| Sl. No | Description | Amended As |
| Pg:45 Para:9.1 10 | Variable focal spot-to-detector between 85cm and 120cm distance and speed up to 9crn/sec or more. | Variable focal spot-to-detector between 90 cm and 119cm distance and speed up to 9crn/sec or more. |
| Pg:45 Para:9.2 9 | Table height 78cm to 104 cm | Table height moves between 80 to 100 cm |
| Pg: 46 Para: 9.4 2 | Anode heat storage capacity should be 2.4 MHU or more having liquid bearing technology or metal lubricant or equivalent performance. The system should have adequate cooling facility for the X-ray tubes for uninterrupted, performance during procedure | Anode heat storage capacity should be 3.0 MHU or more . The Tube should have secondary Grid switch technology or equivalent technology to reduce scattered radiations. The system should have adequate cooling facility for the X-ray tubes for uninterrupted, performance during procedure |
| Pg: 46 Para: 9.6 8 | Nyquist frequency 2.5 lp/mm or higher | Nyquist frequency 2.5 lp/mm or higher |
| pg: 54, Point :16 | OTHER OPTIONAL FEATURES : THE COMPANIES HAVING MODELS WITH ADDITIONAL DOSE SAVING FEATURES MEETING THE FOLLOWING SPECIFICATIONS BELOW SHOULD QUOTE ALTERNATE MODEL. | All additional dose saving measures available with quoted model should be quoted as standard. THE FOLLOWING ITEMS (16 (A,B,C & D) WITH SPECIFICATIONS BELOW SHOULD BE QUOTE AS OPTIONAL |
| | | NOTE: CIVIL FINISHES Sl.no.1 & 2 , Changes updated in Turnkey Breakup in Price Schedule. |



श्री चित्रा तिरुनाल आयुर्विज्ञान और प्रौद्योगिकी संस्थान, तिरुवनंतपुरम- 11, केरल
Sree Chitra Tirunal Institute for Medical Sciences and Technology, Thiruvananthapuram- 11, Kerala
(भारत सरकार के अधीन राष्ट्रीय महत्व संस्थान)
(An Institute of National Importance under Government of India)
टेलीफोन नं./Telephone No. 0471-2443152 फाक्स/Fax 0471-2446433 / 2550728
ई-मेल/E-mail :sct@sctimst.ac.in वेबसाइट/ Website : www.sctimst.ac.in

**SCHEDULE XXI
ELECTRO SURGICAL UNIT**

| SI No | Description | AMENDED AS |
|-------------------------|---|--|
| Corrigendum I Pg: 32 | NEWLY ADDED: 1. UNIVERSAL RETURN ELECTRODE GEL MAT - 5nos | 1. UNIVERSAL RETURN ELECTRODE GEL MAT - 5nos (Optional) |

Equipment Name: Applicable to Schedule No. I to XXIV

| Sl. No | Description | AMENDED AS |
|---------------------------------|---|--|
| In corrigendum dtd 4/12/2019 | EQUIPMENT NOT MANUFACTURED IN INDIA SHOULD BE USFDA 510K CLEARED/ EUROPEAN CE CERTIFIED. ALL OTHERS SHOULD BE BIS/ CDSCO/ AERB APPROVED | EQUIPMENT NOT MANUFACTURED IN INDIA SHOULD BE USFDA 510K CLEARED/ EUROPEAN CE CERTIFIED. ALL OTHERS SHOULD BE BIS/ CDSCO/ AERB APPROVED. WHEREVER INDIAN STANDARDS ARE NOT AVAILABLE THE EQUIPMENT SHOULD HAVE VALID USFDA / EUROPEAN CE CERTIFICATION . |

Note: Revised Technical Compliance and Compliance of BOQ for Turnkey for Schedule -I are provided in excel format in e-tender portal.

All other terms and conditions shall be as per our e-Tender document dtd. 12.11.2019 and subsequent amendments.

Sd/-

DIRECTOR