

CORRIGENDUM

NEURO NAVIGATION SYSTEM FOR CRANIAL & SPINAL APPLICATION (TENDER NO:SCT/004/IMP/2017-18/GJ)

EXISTING	TO BE READ AS
<ul style="list-style-type: none">• System should have facility of virtual fluoroscopic navigation for spinal applications compatible with O-Arm (complete multidimensional surgical imaging system), compatible with intra-operative CT / MRI, Compatible with 9” and 12 “ C-arm of Philips, Shimadzu, Toshiba, Siemens, GE, Ziemhe.	<ul style="list-style-type: none">• System should have facility of virtual fluoroscopic navigation for spinal applications compatible with complete multidimensional surgical imaging systems, compatible with intra-operative CT / MRI, Compatible with 9” and 12 “ C-arm of Philips, Shimadzu, Toshiba, Siemens, GE, Ziemhe.