

## Bid Document

| Bid Details  |  |
|--|--|
| <b>Bid End Date/Time</b>   | 01-12-2022 15:00:00  |
| <b>Bid Opening Date/Time</b>   | 01-12-2022 15:30:00  |
| <b>Bid Offer Validity (From End Date)</b>                                    | 180 (Days)   |
| <b>Ministry/State Name</b>   | Ministry Of Science And Technology   |
| <b>Department Name</b>   | Department Of Science And Technology (dst)   |
| <b>Organisation Name</b>   | Sree Chitra Tirunal Institute For Medical Sciences And Technology (sctimst)  |
| <b>Office Name</b>   | Thiruvananthapuram   |
| <b>Total Quantity</b>  | 45500  |
| <b>Item Category</b>   | intravenous cannulas as per IS: 10555-5 (Q2)   |
| <b>MSE Exemption for Years Of Experience and Turnover</b>                    | Yes  |
| <b>Startup Exemption for Years Of Experience and Turnover</b>                | Yes  |
| <b>Document required from seller</b>   | OEM Authorization Certificate, Compliance of BoQ specification and supporting document<br>*In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| <b>Bid to RA enabled</b>   | No   |
| <b>Primary product category</b>  | intravenous cannulas as per IS: 10555-5  |
| <b>Time allowed for Technical Clarifications during technical evaluation</b> | 2 Days   |
| <b>Evaluation Method</b>   | Item wise evaluation   |

### EMD Detail

|          |    |
|----------|----|
| Required | No |
|----------|----|

### ePBG Detail

|          |    |
|----------|----|
| Required | No |
|----------|----|

(a). The EMD Amount will be applicable for each schedule/group selected during Bid creation.

## Splitting

Bid splitting not applied.

## MII Purchase Preference

|                         |     |
|-------------------------|-----|
| MII Purchase Preference | Yes |
|-------------------------|-----|

## MSE Purchase Preference

|                         |     |
|-------------------------|-----|
| MSE Purchase Preference | Yes |
|-------------------------|-----|

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .In case Buyer has selected Purchase preference to Micro and Small Enterprises clause in the bid, the same will get precedence over this clause.
4. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer)of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY.
5. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

## Evaluation Method ( Item Wise Evaluation Method )

Contract will be awarded schedulewise and the determination of L1 will be done separately for each schedule. The details of item-consignee combination covered under each schedule are as under:

| Evaluation Schedules | Item/Category                           | Quantity |
|----------------------|---|----------|
| Schedule 1           | Intravenous Cannulas As Per Is: 10555-5 | 5000     |
| Schedule 2           | Intravenous Cannulas As Per Is: 10555-5 | 4000     |
| Schedule 3           | Intravenous Cannulas As Per Is: 10555-5 | 21000    |
| Schedule 4           | Intravenous Cannulas As Per Is: 10555-5 | 10000    |
| Schedule 5           | Intravenous Cannulas As Per Is: 10555-5 | 3000     |
| Schedule 6           | Intravenous Cannulas As Per Is: 10555-5 | 2500     |

### Intravenous Cannulas As Per IS: 10555-5 ( 5000 pieces )

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

|            |                  |
|------------|------------------|
| Brand Type | Registered Brand |
|------------|------------------|

### Technical Specifications

[\\* As per GeM Category Specification](#)

| Specification                       | Specification Name  | Bid Requirement (Allowed Values) |
|-------------------------------------|---|----------------------------------|
| Performance Parameters              | <b>Conforming standards for the I V Canulla</b>   | IS / ISO 10555-5                 |
|                                     | <b>Whether IV Canulla is with safety features for preventing needle stick injuries</b>  | Yes                              |
|                                     | <b>Needle Point Finish</b>  | Short bevel cut                  |
|                                     | <b>Needle Hub</b>   | Complying with ISO 594-1         |
|                                     | <b>Needle hub fitting with 6% luer Taper</b>  | Yes                              |
|                                     | <b>Injection Port</b>   | Without                          |
|                                     | <b>Wings</b>  | With                             |
|                                     | <b>Luer Lock Plug/Cap</b>   | Yes                              |
|                                     | <b>Radio opaque line feature</b>  | Yes                              |
| Dimensional and Material Parameters | <b>Material of Catheter</b>   | Polyurethane                     |
|                                     | <b>Nominal outside Diameter of catheter tube and color coding as per IS/ISO 10555-5</b> | Medium Grey - 16G                |

**Consignees/Reporting Officer and Quantity**

| S.No. | Consignee/Reporting Officer | Address  | Delivery Schedule (In number of days from contract start days) |                         |                             |
|-------|-----------------------------|--|--|-------------------------|-----------------------------|
|       |                             |  | Quantity   | Delivery to start after | Delivery to be completed by |
| 1     | Shiju V S                   | 695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum | 1000   | 15                      | 40                          |
|       |                             |  | 1000   | 45                      | 60                          |
|       |                             |  | 2000   | 80                      | 90                          |
|       |                             |  | 1000   | 130                     | 150                         |

**Intravenous Cannulas As Per IS: 10555-5 ( 4000 pieces )**

**(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)**

|            |                  |
|------------|------------------|
| Brand Type | Registered Brand |
|------------|------------------|

**Technical Specifications**

[\\* As per GeM Category Specification](#)

| Specification                       | Specification Name  | Bid Requirement (Allowed Values)     |
|-------------------------------------|---|--------------------------------------|
| Performance Parameters              | <b>Conforming standards for the I V Canulla</b>   | IS / ISO 10555-5                     |
|                                     | <b>Whether IV Canulla is with safety features for preventing needle stick injuries</b>  | Yes                                  |
|                                     | <b>Needle Point Finish</b>  | Long bevel cut                       |
|                                     | <b>Needle Hub</b>   | Complying with ISO 594-1             |
|                                     | <b>Needle hub fitting with 6% luer Taper</b>  | Yes                                  |
|                                     | <b>Injection Port</b>   | Without                              |
|                                     | <b>Wings</b>  | With                                 |
|                                     | <b>Luer Lock Plug/Cap</b>   | Yes                                  |
|                                     | <b>Radio opaque line feature</b>  | Yes                                  |
| Dimensional and Material Parameters | <b>Material of Catheter</b>   | Flourinated Ethylene Propylene (FEP) |
|                                     | <b>Nominal outside Diameter of catheter tube and color coding as per IS/ISO 10555-5</b> | Deep Green - 18G                     |

**Consignees/Reporting Officer and Quantity**

| S.No. | Consignee/Reporting Officer | Address  | Delivery Schedule (In number of days from contract start days) |                         |                             |
|-------|-----------------------------|--|--|-------------------------|-----------------------------|
|       |                             |  | Quantity   | Delivery to start after | Delivery to be completed by |
| 1     | Shiju V S                   | 695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum |  |                         |                             |
|       |                             |  | 1000   | 15                      | 30                          |
|       |                             |  | 1000   | 45                      | 60                          |
|       |                             |  | 2000   | 80                      | 90                          |

**Intravenous Cannulas As Per IS: 10555-5 ( 21000 pieces )**

**(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)**

|            |                  |
|------------|------------------|
| Brand Type | Registered Brand |
|------------|------------------|

**Technical Specifications**

\* As per GeM Category Specification

| Specification                       | Specification Name   | Bid Requirement (Allowed Values) |
|-------------------------------------|--|----------------------------------|
| Performance Parameters              | <b>Conforming standards for the I V Canulla</b>  | IS / ISO 10555-5                 |
|                                     | <b>Whether IV Canulla is with safety features for preventing needle stick injuries</b> | Yes                              |
|                                     | <b>Needle Point Finish</b>   | Short bevel cut                  |
|                                     | <b>Needle Hub</b>  | Complying with ISO 594-1         |
|                                     | <b>Needle hub fitting with 6% luer Taper</b>   | Yes                              |
|                                     | <b>Injection Port</b>  | Without                          |
|                                     | <b>Wings</b>   | With                             |
|                                     | <b>Luer Lock Plug/Cap</b>  | Yes                              |
|                                     | <b>Radio opaque line feature</b>   | Yes                              |
| Dimensional and Material Parameters | <b>Material of Catheter</b>  | Polyurethane                     |

| Specification | Specification Name   | Bid Requirement (Allowed Values) |
|---------------|--|----------------------------------|
|               | Nominal outside Diameter of catheter tube and color coding as per IS/ISO 10555-5 | Pink - 20G                       |

#### Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address  | Delivery Schedule (In number of days from contract start days) |                         |                             |
|-------|-----------------------------|--|--|-------------------------|-----------------------------|
|       |                             |  | Quantity   | Delivery to start after | Delivery to be completed by |
| 1     | Shiju V S                   | 695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum | 4000   | 15                      | 30                          |
|       |                             |  | 4000   | 45                      | 60                          |
|       |                             |  | 10000  | 80                      | 90                          |
|       |                             |  | 3000   | 150                     | 160                         |

#### Intravenous Cannulas As Per IS: 10555-5 ( 10000 pieces )

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

|            |                  |
|------------|------------------|
| Brand Type | Registered Brand |
|------------|------------------|

#### Technical Specifications

[\\* As per GeM Category Specification](#)

| Specification          | Specification Name  | Bid Requirement (Allowed Values) |
|------------------------|---|----------------------------------|
| Performance Parameters | Conforming standards for the I V Canulla  | IS / ISO 10555-5                 |
|                        | Whether IV Canulla is with safety features for preventing needle stick injuries | Yes                              |
|                        | Needle Point Finish   | Short bevel cut                  |
|                        | Needle Hub  | Complying with ISO 594-1         |
|                        | Needle hub fitting with 6% luer Taper   | Yes                              |

| Specification                       | Specification Name  | Bid Requirement (Allowed Values) |
|-------------------------------------|---|----------------------------------|
|                                     | <b>Injection Port</b>   | Without                          |
|                                     | <b>Wings</b>  | With                             |
|                                     | <b>Luer Lock Plug/Cap</b>   | Yes                              |
|                                     | <b>Radio opaque line feature</b>  | Yes                              |
| Dimensional and Material Parameters | <b>Material of Catheter</b>   | Polyurethane                     |
|                                     | <b>Nominal outside Diameter of catheter tube and color coding as per IS/ISO 10555-5</b> | Deep Blue - 22G                  |

#### Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address  | Delivery Schedule (In number of days from contract start days) |                         |                             |
|-------|-----------------------------|--|--|-------------------------|-----------------------------|
|       |                             |  | Quantity   | Delivery to start after | Delivery to be completed by |
| 1     | Shiju V S                   | 695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum | 2000   | 15                      | 30                          |
|       |                             |  | 2500   | 45                      | 60                          |
|       |                             |  | 2500   | 90                      | 100                         |
|       |                             |  | 1500   | 135                     | 150                         |
|       |                             |  | 1500   | 150                     | 180                         |

#### Intravenous Cannulas As Per IS: 10555-5 ( 3000 pieces )

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

|            |                    |
|------------|--------------------|
| Brand Type | Unregistered Brand |
|------------|--------------------|

#### Technical Specifications

[\\* As per GeM Category Specification](#)

| Specification          | Specification Name                              | Bid Requirement (Allowed Values) |
|------------------------|---|----------------------------------|
| Performance Parameters | <b>Conforming standards for the I V Canulla</b> | IS / ISO 10555-5                 |

| Specification                       | Specification Name  | Bid Requirement (Allowed Values)     |
|-------------------------------------|---|--------------------------------------|
|                                     | <b>Whether IV Canulla is with safety features for preventing needle stick injuries</b>  | Yes                                  |
|                                     | <b>Needle Point Finish</b>  | Short bevel cut                      |
|                                     | <b>Needle Hub</b>   | Complying with ISO 594-1             |
|                                     | <b>Needle hub fitting with 6% luer Taper</b>  | Yes                                  |
|                                     | <b>Injection Port</b>   | Without                              |
|                                     | <b>Wings</b>  | With                                 |
|                                     | <b>Luer Lock Plug/Cap</b>   | Yes                                  |
|                                     | <b>Radio opaque line feature</b>  | Yes                                  |
| Dimensional and Material Parameters | <b>Material of Catheter</b>   | Flourinated Ethylene Propylene (FEP) |
|                                     | <b>Nominal outside Diameter of catheter tube and color coding as per IS/ISO 10555-5</b> | Yellow - 24G                         |

#### Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address  | Delivery Schedule (In number of days from contract start days) |                         |                             |
|-------|-----------------------------|--|--|-------------------------|-----------------------------|
|       |                             |  | Quantity   | Delivery to start after | Delivery to be completed by |
| 1     | Shiju V S                   | 695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum |  |                         |                             |
|       |                             |  | 1000   | 15                      | 30                          |
|       |                             |  | 1000   | 45                      | 60                          |
|       |                             |  | 1000   | 110                     | 130                         |

#### Intravenous Cannulas As Per IS: 10555-5 ( 2500 pieces )

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

|            |                  |
|------------|------------------|
| Brand Type | Registered Brand |
|------------|------------------|

#### Technical Specifications

\* [As per GeM Category Specification](#)



| Specification                       | Specification Name  | Bid Requirement (Allowed Values) |
|-------------------------------------|---|----------------------------------|
| Performance Parameters              | <b>Conforming standards for the I V Canulla</b>   | IS / ISO 10555-5                 |
|                                     | <b>Whether IV Canulla is with safety features for preventing needle stick injuries</b>  | Yes                              |
|                                     | <b>Needle Point Finish</b>  | Short bevel cut                  |
|                                     | <b>Needle Hub</b>   | Complying with ISO 594-1         |
|                                     | <b>Needle hub fitting with 6% luer Taper</b>  | Yes                              |
|                                     | <b>Injection Port</b>   | Without                          |
|                                     | <b>Wings</b>  | With                             |
|                                     | <b>Luer Lock Plug/Cap</b>   | Yes                              |
|                                     | <b>Radio opaque line feature</b>  | Yes                              |
| Dimensional and Material Parameters | <b>Material of Catheter</b>   | Polyurethane                     |
|                                     | <b>Nominal outside Diameter of catheter tube and color coding as per IS/ISO 10555-5</b> | Violet - 26G                     |

#### Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address  | Delivery Schedule (In number of days from contract start days) |                         |                             |
|-------|-----------------------------|--|--|-------------------------|-----------------------------|
|       |                             |  | Quantity   | Delivery to start after | Delivery to be completed by |
| 1     | Shiju V S                   | 695011,Sree Chitra Tirunal Institute For Medical Science and Technology,Medical College PO, Trivandrum | 500  | 15                      | 30                          |
|       |                             |  | 1000   | 60                      | 70                          |
|       |                             |  | 1000   | 110                     | 130                         |

#### Special terms and conditions-Version:2 effective from 20-10-2022-Version:1 effective from 04-05-2020 for category intravenous cannulas as per IS: 10555-5

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules made there under as amended till date will always be applicable. This will include all notifications issued by *Central Drugs Standard Control Organisation (CDSCO)*, Ministry of Health & Family Welfare (MoHFW) and Department of

Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.

2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
  3. The price offered by the seller shall not, in any case exceed the DPCO controlled price, if any, fixed by the Central/State Government, the Maximum Retail Price (MRP) and the selling price. The seller must reduce the prices if there is any reduction in DPCO ceiling price, if any.
  4. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e., ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.
2. Special Terms and Conditions for Medical Devices and Consumables covered under Provisions of Drug and Cosmetic Act
1. For items wherever Drug Licence requirements are applicable all provisions of Drug and Cosmetic Act 1940 as amended up to date and Rules made there under will be applicable in addition to any other terms and conditions specified in the Portal.
  2. Drug License: For indigenous products offered in the market, Manufacturer should have valid Drug License as per Drugs and Cosmetic Act 1940 issued by concerned State Drug Control authorities .The Seller if different from the manufacturer shall also be required to be holding Drug License for sale . In case of imported products Manufacturer shall be registered under Form no 10 with Central Drug Authorities ( CDSCO) and the Seller offering imported products should be also holding valid Sales License issued by the local drug authorities. For imported products, certificate from the OEM that product is being used in the Country of Origin should be available with the Seller. It shall be the responsibility of the Seller to ensure that that the Drug License is valid for the product offered and due to any reason the drug control authorities have cancelled or suspended Drug License or convicted the manufacturer or Seller for any offence under the provisions of Drug and Cosmetic Act, Seller should immediately withdraw the product and also intimate the Buyers in case of pending orders for supplies as well as the GeM administration regarding the matter.
  3. Manufacturing & Marketing Experience: Sellers offering the Products in the Portal either as Manufacturers or as Authorised Seller shall ensure that the Products offered are being Manufactured and Marketed in the country (for Indigenous Products) and Marketed (for Imported Products) continuously at least for the last 2 years
  4. Certifications: Manufacturers of offered product (Offered by Manufacturers or by Authorized Seller ) should be holding valid Good Manufacturing Practices Certificate (GMP) as per revised Schedule-M of Drug and Cosmetic Act 1940 as amended up to date or WHO-GMP as per norms amended up to date issued by the Licensing Authority or certificate which is at par with WHO-GMP issued by the authorities of exporting countries / COPP certificate .
  5. Non Conviction Certificate: Sellers either Manufacturers or Authorized Sellers are required to ensure that they are not under conviction in terms of the provisions of Drugs & Cosmetic Act and any other law applicable in relation to the same . In case at any point of time, the Manufacturer or Authorized Seller is convicted under provisions of Drug and Cosmetic Act, it shall be their responsibility to withdraw the product immediately from the market.
  6. Banning and Blacklisting: Seller either Manufacturer or Authorized Seller shall ensure that there is no banning or black listing applicable against them for the product offered on the portal due to quality failure and /or fraudulent/illegal practices or for any other reasons
  7. It shall be the responsibility of the Seller either Manufacturer or Authorized Seller to ensure that manufacturer is having own in-house testing lab to carry out all the required tests as per specification and provisions of drug act as amended up to date for the quoted product and shall also forward the copies of the in-house test reports for each batch along with the supplies. For imported products , certificate from OEM regarding availability of all test facilities in house with them should be available with Seller .
  8. Each lot of supplies shall be dispatched under Self Certification scheme duly supported by in house test reports. Consignees shall be at liberty to draw control Samples and send it to approved Laboratories for testing and in case of any failure , entire responsibility shall rest with Seller in addition to any penalties under the provisions of Drug Act including removal of Goods from the Consignee place. .Further administrative actions as per terms and conditions Gem Portal shall also be applicable.
  9. Packing shall be as per relevant clause of Standard Specifications applicable as indicated in the Catalogue Parameters indicated in the Portal and as per provisions of Drug & Cosmetic Act as amended up to date.
  10. Expiry Date: All supplies must indicate the Month of Manufacture and Expiry. In addition all supplies

- shall have a remaining shelf life of at least 5/6th of the stipulated shelf life at the time of delivery.
11. **Recalls:** If any batch is to be recalled because of problems with product quality or adverse reaction Seller will be responsible to notify the Buyer full details about the reason leading to the recall and shall take steps to replace the product in question at its own cost with a fresh batch of acceptable quality or give a refund of the value of the goods

## **Buyer Added Bid Specific Terms and Conditions**

### **1. Generic**

**OPTION CLAUSE:** The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

### **2. Generic**

**Bidder financial standing:** The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

### **3. Generic**

Bidders shall quote only those products in the bid which are not obsolete in the market and has at least 1 years residual market life i.e. the offered product shall not be declared end-of-life by the OEM before this period.

### **4. Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

### **5. Generic**

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

### **6. Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

Sample (1 no) may be submitted to the office of Sr. Purchase & Stores Officer , Purchase Department, AMCHSS building , SCTIMST on or before 01.12.2022.

### **7. Purchase Preference (Centre)**

Indian suppliers of this item are not allowed to participate and/ or compete in procurement by some foreign governments. Bidders / products from such countries are not eligible / not allowed to participate in this bid in terms of clause 1 (d) of Public Procurement (Preference to Make in India) Order, 2017

### **8. Purchase Preference (Centre)**

Procurement under this bid is reserved for purchase from Micro and Small Enterprises whose credentials are validated online through UDYAM Registration /Udyog Aadhaar (as validated by Government from time to time) for that product category. If the bidder wants to avail the reservation benefit, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after

evaluation of documents submitted.

#### 9. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of Director, SCTIMST A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.

#### 10. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of Director, SCTIMST payable at Thiruvananthapuram. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

#### 11. **Generic**

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

#### 12. **Generic**

**Upload Manufacturer authorization:** Wherever Authorised Distributors are submitting the bid, Manufacturers Authorisation Form (MAF)/Certificate with OEM details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid.

#### 13. **Generic**

Without prejudice to Buyer's right to price adjustment by way of discount or any other right or remedy available to Buyer, Buyer may terminate the Contract or any part thereof by a written notice to the Seller, if:

- i) The Seller fails to comply with any material term of the Contract.
- ii) The Seller informs Buyer of its inability to deliver the Material(s) or any part thereof within the stipulated Delivery Period or such inability otherwise becomes apparent.
- iii) The Seller fails to deliver the Material(s) or any part thereof within the stipulated Delivery Period and/or to replace/rectify any rejected or defective Material(s) promptly.
- iv) The Seller becomes bankrupt or goes into liquidation.
- v) The Seller makes a general assignment for the benefit of creditors.
- vi) A receiver is appointed for any substantial property owned by the Seller.
- vii) The Seller has misrepresented to Buyer, acting on which misrepresentation Buyer has placed the Purchase Order on the Seller.

#### 14. **Generic**

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

## **Disclaimer**

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization. Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity/restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and/or terms and conditions governing the bid. Any clause incorporated by the Buyer such as demanding Tender Sample, incorporating any clause against the MSME policy and Preference to make in India Policy, mandating any Brand names or Foreign Certification, changing the default time period for Acceptance of material or payment timeline

governed by OM of Department of Expenditure shall be null and void and would not be considered part of bid. Further any reference of conditions published on any external site or reference to external documents/clauses shall also be null and void. If any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations. Also, GeM does not permit collection of Tender fee / Auction fee in case of Bids / Forward Auction as the case may be. Any stipulation by the Buyer seeking payment of Tender Fee / Auction fee through ATC clauses would be treated as null and void.

[This Bid is also governed by the General Terms and Conditions](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---Thank You---**